



User instructions

PerfectClean® Cleaning solution refill




Hear well every day
PerfectClean

Switching on - fill the PerfectClean® with the cleaning solution (as soon as the light indicates it).


1

Fully remove the refill tab.




2

Position the refill, top down, in the PerfectClean® refill housing.



3


Push the refill gradually and horizontally up to the stops to pierce the caps. Wait for all the liquid to run out.




How to clean the filter - Before using a new refill.

1

Remove the filter cap

It is important to place the PerfectClean® on its front side with the start button (safe mode to prevent any fluid flow). Remove the filter cap by pivoting it to the left to the  position.




2

Clean the filter cap


Option A



For a little dirty filter:
clean it with warm water.



Option B

For a very dirty filter:
clean it with a cleaning tablet
(20 minutes in a glass) and then
rinse it with warm water.






After cleaning it, replace the filter cap in the  position and turn it to the right, to the  position

Special precautions for use

Only use the refill when the PerfectClean® indicates it.
Completely fill the refill at each fill.

Emptying (to travel or transport)

1. Place the PerfectClean® on its front face (the side with the start button).
2. Remove the filter cap by pivoting it to the left to the  position.
3. Put the PerfectClean® flat over a sink to drain the remaining liquid.
4. Refit the cap in the  position and then turn it to the right to the  position.

Care advice

It is strongly recommended to clean hearing aids every day.
The refill has a use period of about 60 cycles (except the first refill which allows about 30 cycles, which is the minimum operating volume).

Applicable standard

The solution's specific formulation is compliant with the requirements of the NF EN 1040 bactericide standard.

Safety instructions

- Never use a substitute product or liquid for the product present in the cleaning refill.
- The manufacturer cannot be held liable for damage caused by incorrect use or handling.
- Failure to follow these instructions voids the guarantee.

Precautions for use

External use. Do not swallow. Keep out of reach of children.

30 ml e

